

**Table 1****Analysis type: water elements, total**[download ASCII version of this file \(comma separated\)](#) | [download help](#) | [comments?](#)[RMP Data](#) | << [back](#)

Cruise #	Parameter	Units	MDL Target	MDL Measured	Precision Target (+/- %)	Precision Measured (% rpd/rsd) <sup>1</sup>	Accuracy Target (+/- %)	Accuracy Measured (+/- %)	No. Blanks/ Batch
1994-01	Ag	ng/L	1.16	1.01	15	6	25	NA <sup>2</sup>	12/24
1994-01	As	Âµg/L	0.002	0.051	25	5	25	4	2/20
1994-01	Cd	ng/L	0.38	0.3	15	3	25	21	12/24
1994-01	Cr	ng/L	353	9180	15	8	40	27	6/24
1994-01	Cu	ng/L	6.58	5	15	6	25	4	12/24
1994-01	Hg	ng/L	0.1	0.22	25	4	25	5	2/20
1994-01	Ni	ng/L	9.5	12	15	8	25	20	12/24
1994-01	Pb	ng/L	4.97	4	15	2	25	16	12/24
1994-01	Se	Âµg/L	0.005	0.036	35	7	35	7	2/20
1994-01	Zn	ng/L	7.38	4	15	3	25	12	12/24
1994-04	Ag	ng/L	1.16	1.1	15	5	25	NA <sup>2</sup>	8/24
1994-04	As	Âµg/L	0.002	0.068	25	7	25	3	2/20
1994-04	Cd	ng/L	0.38	0.429	15	1	25	19	8/24
1994-04	Cr	ng/L	42	0.088	15	9	40	35	5/24
1994-04	Cu	Âµg/L	0.007	0.018	15	4	25	15	8/24
1994-04	Hg	ng/L	0.1	0.14	25	3	25	3	2/20
1994-04	Ni	Âµg/L	0.009	0.026	15	3	25	23	8/24
1994-04	Pb	Âµg/L	0.003	0.002	15	7	25	25	8/24
1994-04	Se	Âµg/L	0.005	0.018	35	11	35	7	2/20
1994-04	Zn	Âµg/L	0.005	0.021	15	4	25	6	8/24
1994-08	Ag	ng/L	1.16	0.664	15	5	25	NA <sup>2</sup>	12/24
1994-08	As	Âµg/L	0.002	0.09	25	10	25	10	2/20
1994-08	Cd	Âµg/L	0.38	0.0004	15	5	25	NA <sup>3</sup>	12/24
1994-08	Cr	ng/L	353	337	15	13	40	28	
1994-08	Cu	Âµg/L	6.58	0.018	15	4	25	23	12/24
1994-08	Hg	ng/L	0.1	0.44	25	10	25	4	2/20
1994-08	Ni	Âµg/L	9.5	8	15	7	25	15	
1994-08	Pb	Âµg/L	4.97	7	15	7	25	5	
1994-08	Se	Âµg/L	0.005	0.051	35	12	35	7	2/20
1994-08	Zn	Âµg/L	7.38	0.004	15	4	25	5	11/24

<sup>1</sup> relative percent difference or relative standard deviation

<sup>2</sup> There are no SRM certified values for silver.

<sup>3</sup> Not available

<sup>4</sup> Not available due to laboratory error.